

### Basic data

Style: 4711  
Reference: ASTM D737  
Date: 19.04.2012  
Time: 10:50:24  
Instrument: FX 3300 LabAir IV  
Serial Number: 144


### Settings

Test pressure: 125 Pa  
Test area: 38 cm<sup>2</sup>  
Nom / Min / Max: 4.00 / 3.70 / 4.30 ft<sup>3</sup>/ft<sup>2</sup>/min

### Statistical analysis

Average: 4.00 ft<sup>3</sup>/ft<sup>2</sup>/min  
Minimum: 3.68 ft<sup>3</sup>/ft<sup>2</sup>/min  
Maximum: 4.22 ft<sup>3</sup>/ft<sup>2</sup>/min  
CV: 4.44 %  
Cpk: 0.560

### Test results



1: 4.22 ft<sup>3</sup>/ft<sup>2</sup>/min  
2: 3.98 ft<sup>3</sup>/ft<sup>2</sup>/min  
3: ▲ 3.68 ft<sup>3</sup>/ft<sup>2</sup>/min  
4: 4.09 ft<sup>3</sup>/ft<sup>2</sup>/min  
5: 4.04 ft<sup>3</sup>/ft<sup>2</sup>/min